

**PRODUCT INFORMATION**

Product SKU	CTNGPOS
Product Name	CT/NG Positive Run Control
Kit Lot Number	22081102
Source Material	<i>Chlamydia trachomatis</i> <i>Neisseria gonorrhoeae</i>
Date of Expiration	2024-01-31
Base Matrix	Synthetic Matrix
Preservative	Sodium Azide
Fill Volume	1.1 mL
Storage Temperature	2-8°C
Regulatory Status	For In Vitro Diagnostic Use
Testing Platform	In-House CT/NG Assay

This is an unassayed control and does not have assigned analyte values. The end user is responsible for determining the value of the control. This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.


As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

**TESTING DATA**

Lot Number	Product Description	Target Concentrations	Testing Results (Comparison of Target Value Against Reference Standard)	Quantity
22081102	CT/NG Positive Run Control	CT: 100,000 cp/mL NG: 100,000 cp/mL	Pass	6 Vials

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

**APPROVAL**

Quality Assurance	Date
	03 Oct 22