

**PRODUCT INFORMATION**

Product SKU	EBVP100
Product Name	EBV Verification Panel
Kit Lot Number	21101806
Source Material	Inactivated EBV
Date of Expiration	2024-09-30
Base Matrix	True EDTA Plasma
Preservative	ProClin® 300
Fill Volume	1.0 mL
Storage Temperature	-20°C or Below
Regulatory Status	CE-IVD
Traceability	1 <sup>st</sup> WHO International Standard (09/260)
Testing Platform	In-House EBV Assay

This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.

As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.


**TESTING DATA**

Lot Number	Product Description	Target Concentrations (IU/mL)*	Results
21101801	EBV Panel Member 1: 5E2 IU/mL	500 IU/mL	Pass
21101802	EBV Panel Member 2: 5E3 IU/mL	5,000 IU/mL	Pass
21101803	EBV Panel Member 3: 5E4 IU/mL	50,000 IU/mL	Pass
21101804	EBV Panel Member 4: 5E5 IU/mL	500,000 IU/mL	Pass
21101805	EBV Panel Member 5: 5E6 IU/mL	5,000,000 IU/mL	Pass

\*1 International Unit = 0.49 digital PCR copy

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

**APPROVAL**

Quality Assurance	Date
	26 Oct 21