

**PRODUCT INFORMATION**

Product SKU	HCVGTP
Product Name	HCV Genotype Panel
Kit Lot Number	64508810
Source Material	Inactivated HCV
Date of Expiration	2024-02-29
Base Matrix	True EDTA Plasma
Preservative	ProClin® 300
Fill Volume	1.0 mL
Storage Temperature	-20°C or Below
Regulatory Status	For Research Use Only. Not for use in diagnostic procedures.
Traceability	4 <sup>th</sup> WHO International Standard (06/102)
Testing Platform	Roche COBAS® AmpliPrep/COBAS® TaqMan® HCV v2.0

This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.

As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

**TESTING DATA**

Lot Number	Product Description	Target Concentration	Results
64508810-1	HCV Genotype 1a	20,000 IU/mL	Pass
64508810-2	HCV Genotype 1b	20,000 IU/mL	Pass
64508810-3	HCV Genotype 2	20,000 IU/mL	Pass
64508810-4	HCV Genotype 3	20,000 IU/mL	Pass
64508810-5	HCV Genotype 4	20,000 IU/mL	Pass
64508810-6	HCV Genotype 5	20,000 IU/mL	Pass

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

**APPROVAL**

Quality Assurance	Date
<i>Melody Huang</i>	<i>16 Sep 22</i>