

**PRODUCT INFORMATION**

<b>Product SKU</b>	HCVH102
<b>Product Name</b>	HCV High Run Control
<b>Kit Lot Number</b>	21083006
<b>Source Material</b>	Recombinant HCV
<b>Date of Manufacture</b>	2021-09-09
<b>Date of Expiration</b>	2024-08-31
<b>Base Matrix</b>	True EDTA Plasma
<b>Preservative</b>	ProClin® 300
<b>Fill Volume</b>	1.0 mL
<b>Storage Temperature</b>	-20°C or below
<b>Intended Use</b>	CE-IVD
<b>Traceability</b>	4 <sup>th</sup> HCV International Standard (06/102)
<b>Testing Platform</b>	In-House HCV Assay

This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.


As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

**TESTING DATA**

Lot Number	Product Description	Target Concentration	Testing Results	Quantity
21083006	HCV High Run Control	1,000,000 IU/mL	Pass (Comparison of target value against reference standard)	5 Vials

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

**APPROVAL**

Quality Assurance	Date
	09 Aug 22