

**PRODUCT INFORMATION**

<b>Product SKU</b>	HPVP200
<b>Product Name</b>	HPV Verification Panel
<b>Kit Lot Number</b>	21081009
<b>Source Material 1</b>	HPV Genotype 16 Cell Line
<b>Source Material 2</b>	HPV Genotype 18 Cell Line
<b>Source Material 3</b>	HPV Genotype 68 Cell Line
<b>Source Material 4</b>	HPV Negative Cell Line
<b>Date of Expiration</b>	2024-07-31
<b>Base Matrix</b>	PreservCyt®
<b>Fill Volume</b>	1.4 mL
<b>Storage Temperature</b>	2-8°C
<b>Regulatory Status</b>	For In Vitro Diagnostic Use
<b>Testing Platform</b>	cobas® 6800/8800 HPV Test and In-House HPV Negative Assay


This is an unassayed control and does not have assigned analyte values. The end user is responsible for determining the value of the control. This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.

As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

**TESTING DATA**

Lot Number	Product Description	Target Concentration	Quantity	Results
21081010	HPV 16 Genotype 3x LOD	HPV-16: 60 cp/mL	1 vial	Pass
		Beta-Globin: 150,000 cp/mL		
21081011	HPV 16 Genotype 5x LOD	HPV-16: 100 cp/mL	1 vial	Pass
		Beta-Globin: 150,000 cp/mL		
21081012	HPV 16 Genotype 10x LOD	HPV-16: 200 cp/mL	1 vial	Pass
		Beta-Globin: 150,000 cp/mL		
21081013	HPV 18 Genotype 3x LOD	HPV-18: 60 cp/mL	1 vial	Pass
		Beta-Globin: 150,000 cp/mL		
21081014	HPV 18 Genotype 5x LOD	HPV-18: 100 cp/mL	1 vial	Pass
		Beta-Globin: 150,000 cp/mL		
21081015	HPV 18 Genotype 10x LOD	HPV-18: 200 cp/mL	1 vial	Pass
		Beta-Globin: 150,000 cp/mL		
21081016	HPV 16/18/68 Genotype 3x LOD	HPV-16: 60 cp/mL	3 vial	Pass
		HPV-18: 60 cp/mL		
		HPV-68: 600 cp/mL		
		Beta-Globin: 150,000 cp/mL		
21081017	HPV 16/18/68 Genotype 5x LOD	HPV-16: 100 cp/mL	1 vial	Pass
		HPV-18: 100 cp/mL		
		HPV-68: 1,000 cp/mL		
		Beta-Globin: 150,000 cp/mL		
21081018	HPV 16/18/68 Genotype 10x LOD	HPV-16: 200 cp/mL	1 vial	Pass
		HPV-18: 200 cp/mL		
		HPV-68: 2,000 cp/mL		
		Beta-Globin: 150,000 cp/mL		
21081019	HPV Negative Bundle Control	Beta-Globin: 150,000 cp/mL	5 vial	Pass

**APPROVAL**

<b>Quality Assurance</b>	<b>Date</b>
	16sep21