

PRODUCT INFORMATION

Product SKU	HPVP200
Product Name	HPV Verification Panel
Kit Lot Number	22051911
Source Material 1	HPV Genotype 16 Cell Line
Source Material 2	HPV Genotype 18 Cell Line
Source Material 3	HPV Genotype 68 Cell Line
Source Material 4	HPV Negative Cell Line
Date of Expiration	2025-04-30
Base Matrix	PreservCyt®
Fill Volume	1.4 mL
Storage Temperature	2-8°C
Regulatory Status	For In Vitro Diagnostic Use
Testing Platform	cobas® 6800/8800 HPV Test and In-House HPV Negative Assay


This is an unassayed control and does not have assigned analyte values. The end user is responsible for determining the value of the control. This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.

As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

TESTING DATA

Lot Number	Product Description	Target Concentration	Quantity	Results
22051901	HPV 16 Genotype 3x LOD	HPV-16: 60 cp/mL	1 vial	Pass
		Beta-Globin: 150,000 cp/mL		
22051902	HPV 16 Genotype 5x LOD	HPV-16: 100 cp/mL	1 vial	Pass
		Beta-Globin: 150,000 cp/mL		
22051903	HPV 16 Genotype 10x LOD	HPV-16: 200 cp/mL	1 vial	Pass
		Beta-Globin: 150,000 cp/mL		
22051904	HPV 18 Genotype 3x LOD	HPV-18: 60 cp/mL	1 vial	Pass
		Beta-Globin: 150,000 cp/mL		
22051905	HPV 18 Genotype 5x LOD	HPV-18: 100 cp/mL	1 vial	Pass
		Beta-Globin: 150,000 cp/mL		
22051906	HPV 18 Genotype 10x LOD	HPV-18: 200 cp/mL	1 vial	Pass
		Beta-Globin: 150,000 cp/mL		
22051907	HPV 16/18/68 Genotype 3x LOD	HPV-16: 60 cp/mL	3 vials	Pass
		HPV-18: 60 cp/mL		
		HPV-68: 600 cp/mL		
		Beta-Globin: 150,000 cp/mL		
22051908	HPV 16/18/68 Genotype 5x LOD	HPV-16: 100 cp/mL	1 vial	Pass
		HPV-18: 100 cp/mL		
		HPV-68: 1,000 cp/mL		
		Beta-Globin: 150,000 cp/mL		
22051909	HPV 16/18/68 Genotype 10x LOD	HPV-16: 200 cp/mL	1 vial	Pass
		HPV-18: 200 cp/mL		
		HPV-68: 2,000 cp/mL		
		Beta-Globin: 150,000 cp/mL		
22051910	HPV Negative Bundle Control	Beta-Globin: 150,000 cp/mL	5 vials	Pass

APPROVAL

Quality Assurance	Date
	01 June 22