

**PRODUCT INFORMATION**

<b>Product SKU</b>	HSV2P100
<b>Product Name</b>	HSV-2 Verification Panel
<b>Kit Lot Number</b>	64553054
<b>Source Material</b>	Inactivated HSV-2
<b>Date of Expiration</b>	2025-04-30
<b>Base Matrix</b>	Synthetic CSF Matrix
<b>Preservative</b>	ProClin® 300
<b>Fill Volume</b>	1.0 mL
<b>Storage Temperature</b>	-20°C or Below
<b>Regulatory Status</b>	CE-IVD
<b>Testing Platform</b>	In-House HSV-2 Assay

This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.

As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

**TESTING DATA**

Lot Number	Product Description	Target Concentrations	Results
64553054-1	HSV-2 Panel Member 1: 2E2 cp/mL	200 cp/mL	Pass
64553054-2	HSV-2 Panel Member 2: 2E3 cp/mL	2,000 cp/mL	Pass
64553054-3	HSV-2 Panel Member 3: 2E4 cp/mL	20,000 cp/mL	Pass
64553054-4	HSV-2 Panel Member 4: 2E5 cp/mL	200,000 cp/mL	Pass
64553054-5	HSV-2 Panel Member 5: 2E6 cp/mL	2,000,000 cp/mL	Pass

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

**APPROVAL**

Quality Assurance	Date
<i>Melody Harrington</i>	<i>22 May 23</i>