

PRODUCT INFORMATION

Product SKU	NEGSTI
Product Name	STI Negative Run Control
Kit Lot Number	22081101
Source Material	HPV Negative Cell Line
Date of Expiration	2024-01-31
Base Matrix	Synthetic Matrix
Preservative	Sodium Azide
Fill Volume	1.1 mL
Storage Temperature	2-8°C
Regulatory Status	For In Vitro Diagnostic Use
Testing Platform	In-House HPV Assay

This is an unassayed control and does not have assigned analyte values. The end user is responsible for determining the value of the control. This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.

As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

TESTING DATA

Lot Number	Product Description	Target Concentration	Testing Results (Comparison of Target Value Against Reference Standard)	Quantity
22081101	STI Negative Run Control	Beta-Globin: 75,000 cp/mL	Pass	6 Vials

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

APPROVAL

Quality Assurance <i>Melody Hays</i>	Date 05 Oct 22
--	--------------------------