

PRODUCT INFORMATION

Product SKU	VZVP100
Product Name	VZV Verification Panel
Kit Lot Number	21092218
Source Material	Inactivated VZV
Date of Expiration	2023-08-31
Base Matrix	Synthetic CSF Matrix
Preservative	ProClin® 300
Fill Volume	1.0 mL
Storage Temperature	-20°C or Below
Regulatory Status	CE-IVD
Testing Platform	In-House VZV Assay

This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.


As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

TESTING DATA

Lot Number	Product Description	Target Concentrations	Results
21092213	VZV Panel Member 1: 5E2 cp/mL	500 cp/mL	Pass
21092214	VZV Panel Member 2: 2E3 cp/mL	2,000 cp/mL	Pass
21092215	VZV Panel Member 3: 2E4 cp/mL	20,000 cp/mL	Pass
21092216	VZV Panel Member 4: 2E5 cp/mL	200,000 cp/mL	Pass
21092217	VZV Panel Member 5: 2E6 cp/mL	2,000,000 cp/mL	Pass

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

APPROVAL

Quality Assurance 	Date 05Oct21
-----------------------------------------------------------------------------------------------------------------	------------------------