

PRODUCT INFORMATION

Product SKU	ZKVP100
Product Name	Zika Verification Panel
Kit Lot Number	64563426
Source Material	Inactivated Zika Virus
Date of Expiration	2026-06-30
Base Matrix	True EDTA Plasma
Preservative	ProClin® 300
Fill Volume	1.0 mL
Storage Temperature	-70°C or Below
Regulatory Status	CE-IVD
Testing Platform	In-House Zika Assay

This quality control material should be used in accordance with local, state, and/or federal regulations or accreditation requirements.


As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols.

TESTING DATA

Lot Number	Product Description	Target Concentrations	Results
64563426-1	Zika Panel Member 1: 1E2 cp/mL	100 cp/mL	Pass
64563426-2	Zika Panel Member 2: 1E3 cp/mL	1,000 cp/mL	Pass
64563426-3	Zika Panel Member 3: 1E4 cp/mL	10,000 cp/mL	Pass
64563426-4	Zika Panel Member 4: 1E5 cp/mL	100,000 cp/mL	Pass
64563426-5	Zika Panel Member 5: 1E6 cp/mL	1,000,000 cp/mL	Pass

Bioburden Testing Results	
Acceptance Criteria: Less than 10 CFU/mL	Pass

APPROVAL

Quality Assurance	Date
	2025/1/23