

EDX RP Negative Run Control Summary	
Catalog Number	RPNEG
Nonreactive Analytes	<p>Adenovirus  <i>Bordetella parapertussis</i>  <i>Bordetella pertussis</i>  <i>Chlamydia pneumoniae</i>            Coronavirus 229E            Coronavirus HKU1 (RNA IVT)            Coronavirus NL63 (RNA IVT)            Coronavirus OC43            Human Metapneumovirus (RNA IVT)            Influenza A – H1N1            Influenza A – H1N1 2009 (RNA IVT)            Influenza A – H3N2 (RNA IVT)            Influenza B  <i>Mycoplasma pneumoniae</i>            Middle Eastern Respiratory Syndrome (RNA IVT)            Parainfluenza 1            Parainfluenza 2            Parainfluenza 3            Parainfluenza 4            Rhinovirus            Respiratory Syncytial Virus A            Respiratory Syncytial Virus B            Severe Acute Respiratory Syndrome Coronavirus 2</p>
Matrix	Synthetic Matrix
Preservative	ProClin® 300
Storage	-20°C or Below
Fill Volume	0.3mL
Number of Vials Per Kit	6 Vials
Number of Uses	Single Use
Shelf Life	18 Months from Date of Manufacture
Precautions	Biological Source Material. Treat as Potentially Infectious
Regulatory Status	For In Vitro Diagnostic Use

**INTENDED USE:**

The EDX RP Negative Run Control is an external quality control, intended to be used with respiratory assays to monitor the performance of that given assay. Routine use of EDX RP Negative Run Control allows laboratories to meet their individualized quality control plans including evaluation of reagent lot changes and new shipments, monitoring of multiple identical devices, and assessment of different personnel and locations.

**PRODUCT DESCRIPTION:**

The EDX RP Negative Run Control is negative for 23 respiratory analytes. The EDX RP Negative Run Control is formulated in synthetic matrix which contains preservatives.

**PROCEDURE:**

The EDX RP Negative Run Control requires an extraction step. The control must be treated in a similar manner to other tested samples. This allows the operator to assess their extraction technique.

The EDX RP Negative Run Control is an unassayed control. The control is not a substitute for the manufacturer’s kit controls provided with the assay. EDX RP Negative Run Control is to be tested according to the assay manufacturer’s or testing laboratory’s instructions and recommendations. Expected results utilizing EDX RP Negative Run Control must be established by the end user for their specific assay.

**STORAGE AND HANDLING:**

The EDX RP Negative Run Control should be stored at -20°C or below.

Thaw EDX RP Negative Run Control at room temperature and vortex briefly prior to use.

Once thawed, EDX RP Negative Run Control is stable for 24 hours when stored at 2-8°C. Any remaining materials must be disposed of after one use. Do not reuse. Do not dilute.

Do not use EDX RP Negative Run Control beyond the expiration date.

**LIMITATIONS:**

For In Vitro Diagnostic Use.

**WARNINGS AND PRECAUTIONS:**












The EDX RP Negative Run Control should be treated as potentially infectious. Universal precautions and proper disposal should be practiced<sup>1</sup>. Do not pipette by mouth. Do not smoke, eat, or drink in areas where specimens are handled.

Discard product if packaging is damaged or leaking. Disinfect liquids, materials or spills with a 0.5% sodium hypochlorite solution. Dispose of all materials and liquids used in the procedure as if they contained pathogenic agents.

### REFERENCES:

<sup>1</sup> Siegel JD, Rhinehart E, Jackson M, Chiarello L, and the Healthcare Infection Control Practices Advisory Committee, 2007  
Guideline for Isolation Precautions: Preventing Transmission

### SYMBOL REFERENCES:

	Catalog Number
	Lot Number
	For In Vitro Diagnostic Use
	Expiration Date
	Upper Limit of Temperature
	Manufacturer
	Caution
	Consult Instructions for Use
	Negative Control
	European Mark of Conformity
	Authorized Representative



Emergo Europe  
Prinsessegracht 20  
2514 AP The Hague  
The Netherlands



Exact Diagnostics LLC  
100 South Jones Street, Suite 100,  
Fort Worth, Texas 76104  
United States



**Bio-Rad  
Laboratories**

### Clinical Diagnostics Group

9500 Jeronimo Road  
Irvine, California 92618  
(800) 854-6737  
FAX (949) 598-1550  
bio-rad.com/qualitycontrol

**Technical Service:**  
(800) 854-6737

**United States Customer Service:** 4000 Alfred Nobel Drive • Hercules, California 94547 • Phone (510) 724-7000 • FAX (510) 741-6373  
**Also in:** Gladesville, Australia, Phone +61 (2) 9914 2800 • Fax +61 (2) 9914 2888 **Vienna, Austria,** Phone +43 (0) 1 877 89 01 9 • Fax +43 (0) 1 876 56 29 **Temse, Belgium,** Phone +32 (0) 3 710 53 00 • Fax +32 (0) 3 710 53 01 **São Paulo, Brazil,** Phone +55 11 3065-7550 **Montréal, Canada,** Phone +1 514 334 4372 • Fax +1 514 334 0872 **Shanghai, China,** Phone +86 21 6169 8500 • Fax +86 21 6169 8599 **Prague, Czech Republic,** Phone +420 241 431 660 **Symbion Science Park, Denmark,** Phone +45 44 52 10 00 • Fax +45 44 52 10 01 **Helsinki, Finland,** Phone +358 9 804 22 00 **Marnes-la-Coquette, France,** Phone +33 (0)1 47 95 60 00 • Fax +33 (0)1 47 41 91 33 **Munich, Germany,** Phone +49 (0) 89 31884 393 • Fax +49 (0) 89 31884 136 **Athens, Greece,** Phone +30 210 7774396 • Fax +30 210 7774376 **Quarry Bay, Hong Kong,** Phone +85 2 2789 3300 • Fax +85 2 2789 1257 **Budapest, Hungary,** Phone +36 1 459 6190 • Fax +36 1 459 6101 **Haryana, India,** Phone +91 124 4029300 • Fax +91 124 2398115 **Rishon Le Zion, Israel,** Phone +972 03 963 6025 • Fax +972 03 951 4129 **Milan, Italy,** Phone +39 024 94 86 600 • Fax +39 02 21609399 **Tokyo, Japan,** Phone +81 3 6361 7070 • Fax +81 3 5463 8481 **Seoul, Korea,** Phone +82 080 007 7373 • Fax +82 (2) 3472 7003 **Mexico D.F., Mexico,** Phone +52 (55) 5488 7670 • Fax +52 (55) 1107 7246 **Veenendaal, The Netherlands,** Phone +31 (0)318 540 666 • Fax +31 (0)318 542 216 **Auckland, New Zealand,** Phone +64 (9)415 2280 • Fax +64 (9)415 2284 **Oslo, Norway,** Phone +47 23 38 41 30 • Fax +47 23 38 41 39 **Warsaw, Poland,** Phone +48 22 331 99 99 • Fax +48 22 331 99 88 **Amadora, Portugal,** Phone +351 21 47 27 700 • Fax +351 21 472 7777 **Moscow, Russia,** Phone +7 495 721 1404 • Fax +7 495 721 1412 **Singapore,** Phone +65 6415 3170 • Fax +65 6415 3189 **Johannesburg, South Africa,** Phone +27 11 442 8508 • Fax +27 11 442 8525 **Madrid, Spain,** Phone + 34 91 490 6580 • Fax +34 91 590 5211 **Sundbyberg, Sweden,** Phone +46 844 98053 • Fax +46 8 55 51 27 80 **Cressler, Switzerland,** Phone +41 (0) 61 717 9555 • Fax +41 (0) 61 717 9550 **Taipei, Taiwan,** Phone +886 (2) 2578-7189 • Fax +886 (2) 2578-6890 **Bangkok, Thailand,** Phone (662) 651 8311 • Fax (662) 651 8312 **Hertfordshire, United Kingdom,** Phone +44 (0)1923 471301 • Fax +44 (0)1923 471340